



AQU@Sense MB: A technological evaluation for « on-line » pharmaceutical water analysis

Co-authors:

Lucile PLOURDE-OWOBI^{(1)*} and Felix THIELE⁽²⁾

Contributors:

Axel ARTHUR⁽³⁾, Stéphanie BARON⁽¹⁾, Pierre CULLMANN⁽²⁾, Christophe GEORGES⁽²⁾,
Nadège HENNY⁽¹⁾, Mickaël MALLEN⁽¹⁾, David PACORY⁽¹⁾, Patrice RIOU⁽¹⁾ and Christian RUFFE⁽²⁾

(1) Sanofi Pasteur - 69280 Marcy L'étoile *Corresponding author: lucile.pourde@sanofi .com

(2) BWT - Paris / Suisse

(3) INSA - 31400 Toulouse

THE «BIOBURDEN» CONTROL is a release test which allows the microbial load of products to be determined. In accordance with the standards of the Pharmacopoeias, it can be carried out *via* a culture step on a specified medium (R2A plates for water testing) by filtration, surface spreading or by in-depth seeding. The result obtained after 5 regulatory days of incubation at 30–35°C is expressed in CFU (Colony Forming Unit) per volume filtered.

This Pharmacopoeia test has known limitations, such as the long duration before obtaining the result, the high rate of handling errors or a poor recovery of the microflora (<1% cultivable).


Alternative methods were first developed to reduce analysis time and are already in place in the laboratory performing the tests. On the industrial side, real-time analyzers are now required in order to monitor the "bioburden" and manage contamination control.

► **New technologies are now applied to the "monitoring" of water systems:** **AQU@Sense MB** is marketed by the company BWT, the principle of microbial detection is based on DNA labeling and analysis of fluorescence events by **flow cytometry**.

Water is a key element in the pharmaceutical industry because it is used as a raw material, diluent, or even as a final product. There are 2 grades of pharmaceutical water: Purified Water (PW) and Water For Injection (WFI). The microbiological control of water is thus strictly regulated:


Pharmaceutical quality	WFI	PW
Microbial load / 100 ml (bioburden)	< 10 CFU	< 10,000 CFU

The **AQU@Sense MB** counts the actual bacterial cells present in water, accurately & within 20 minutes. Even bacteria which are in a stress-induced state, that cannot be cultivated (called VBNC), are detected by this flow cytometer. Additionally, the method does not depend on incubational and nutritional requirements.




- Automated flow cytometer
- Continuous monitoring of Water
- DNA specific stain (SYBR Green): no false measurement of particles

OPERATING PRINCIPLE:



- 1 SAMPLING**
Integrated automatic sampling and processing saves time in the lab.
- 2 PREPARATION**
The fully automated preparation of the sample includes dyeing, mixing and incubation.
- 3 MEASUREMENT**
A laser induces cell fluorescence. A detector analyses the cells based on their fluorescent response signal. The cell count is as accurate as using lab equipment.
- 4 ANALYSIS**
The results are available in 20 minutes. See total cell count (TCC) and share of HNA/LNA (high and low nucleic acid content).
- 5 VISUALISATION**
The information and values are displayed on the integrated HMI. Data export is extremely user-friendly.
- 6 CLEANING**
The cleaning process is automated. Ready for the next sample.
- 7 CHEMICALS AND SAMPLE**
All generated waste is collected in the sealed cartridge, which contains, at the same time, the chemicals required for ~1.000 measurements.



RESULT ANALYSIS :

TCC = Total Cell Count

FL1 = Fluorescence at 525 nm

FL2 = Fluorescence at 715 nm

Background = SYBR Green, cell fragments or particles

Noise = Electrical noise from the optical detector

The objective of the technological evaluation of AQU@Sense MB is to determine the performance of the system applied for on-line water monitoring 1) by generating data from laboratory assays using known and calibrated microbial strains and 2) by performing measurements and testing different water loops at sample points.

- **Equipment installation and training of personnel** < 1 day
- **Protocol for sampling, testing and analysis according to BWT procedure** (see operating principle) < 30 min



STUDY DESIGN SPACE = Lab assays (microbial testing) and on-line water monitoring of different water loops (production or distribution)

- **Different matrix assessed** (D fluid, TBS buffer & nuclease free water) = compatibility verification (no matrix interference)
- **Spiking assay with known and calibrated strains** (ATCC reference Pharmacopeia strains (500 CFU bioball – bioMérieux) + 2 wild isolates (from Marcy L'étoile, < - 70°C frozen samples)
 - ▶ specificity < 100 CFU = detection of all viable microflora
 - ▶ range & linearity (between 10 to 10⁶ CFU – duplicate assays)
- **Repeatability assay (4 replicates)**
- **Monitoring of PW loops** (production & distribution installation of PW) and « bioburden » testing according to the current method

« **Bioburden** » testing is performed according to the Pharmacopeia method by a qualified person: 100 mL sample is filtered (duplicated assay) using Milliflex cassettes (R2A Agar medium) that are incubated at 30–35°C during at least 5 days. Results are interpreted at the 6th day of incubation by


observation of colonies present on the surface of the membrane: number of CFU (Colony Form Unit) per filtered volume is then determined.

Inoculum control (duplicated assay) is performed by culture on Agar plates (TSA) incubated during 2 days at 30–35°C.



Reference strains	<i>Escherichia coli</i> ATCC 12923, <i>Bacillus subtilis</i> ATCC 10400, <i>Pseudomonas aeruginosa</i> ATCC12924, <i>Staphylococcus aureus</i> ATCC 10788, <i>Aspergillus brasiliensis</i> NCPF 2275 and <i>Candida albicans</i> NCPF 3179
Environmental isolates	<i>Ralstonia picketti</i> and <i>Staphylococcus epidermidis</i> (in-house strains preserved according Sanofi Pasteur procedure)

RESULTS - MICROBIAL EVALUATION (OFF-LINE)

REPEATABILITY RESULTS: germ-free & nuclease-free water




REPEATABILITY RESULTS: *E. COLI* (500 CFU/mL) in nuclease-free water





CFU Colony Forming Unit
≠ TCC Total Cell Count

SPECIFICITY: Examples of spiked samples with known microflora detected by AQU@Sense MB (HPC : TSA, 2days @ 30–35°C)





LIMIT OF QUANTIFICATION:

10⁸ CFU/mL sample is serially diluted 10-fold: the cell concentrations defined between **1 CFU /mL and 10⁶ CFU /mL** are detected quasi-linearly.





RESULTS - INDUSTRIAL EVALUATION (ON-LINE)

**« ON-LINE » RESULTS IN PRODUCTION PW SYSTEM:
WITHOUT REPORTED INCIDENT (R#1)**



**« ON-LINE » RESULTS IN DISTRIBUTION PW SYSTEM:
WITHOUT REPORTED INCIDENT (R#2)**





The TCC data collected continuously showed the importance of the analysis frequencies. Online FCM enables a detailed characterization of the frequency and magnitude of microbiological fluctuations on a level of detail not possible with standard methods.


« ON-LINE » RESULTS IN PRODUCTION PW SYSTEM: with 2 series of reported incidents (R#2)

A TCC baseline is present with some sporadic events that do not directly correlate with variations in the physical parameters of the loop (flow, filling of

the tank, pressure, total organic carbon, or conductivity), except for a reported chlorometer issue with abnormal injected volumes of NaOH (Oct. 30th and Nov. 3rd). The associated events clearly indicate a microbiological problem (see examples on the right).



First example




Second example

► As a first example, the reported 2,240 TCC (red point) is clearly associated with a microbial event (as indicated by the bacterial clusters within the gate). These cells are possibly VBNC, as the plate count showed no elevated number of CFU.

► As a second example, no microbial event is reported = 0 TCC (green point part of the baseline).

ON-LINE MEASUREMENTS WERE PERFORMED DURING DIFFERENT PERIODS OF TIME AND AT DIFFERENT SAMPLE POINTS FROM PW LOOPS OF PRODUCTION OR DISTRIBUTION SYSTEMS:

- **R#1 system** produces a demineralized and osmosis treated PW characterized by 0.04 $\mu\text{S}/\text{cm}$ conductivity, 3 ppb TOC with a variable flow: 40 – 60 m^3/hr .
- **R#2 system** produces EDI and osmosis treated PW characterized by 0.06 $\mu\text{S}/\text{cm}$ conductivity, TOC=1 ppb with a variable flow: 9 – 13 m^3/hr . (EDI: Electrodesionization)
- **R#2 distribution system** is illustrated below:



CONCLUSION

The microbial results have allowed to check the performance of the AQU@Sense MB in terms of microbial detection in high concentration ranges (from 1 to 10^6 CFU/mL), except for mold (*Aspergillus brasiliensis*).

The technological principle remains simple and the device implementation is fast. Alert and action limits are still to be defined during "on-line" monitoring.

Microbial quality of Purified Water is by this technology continuously monitored and controlled to prevent contamination.

This initial industrial assessment has been performed using the first generation of the cartridge (TCC). BWT has recently introduced a new cartridge with an improved chemical formulation, which can directly quantify the number of viable bacteria based on the assessment of bacterial membrane integrity (ICC: Intact Cell Count), and it should be tested further!

© Sanofi and BWT